

ART 4.3.4.1 Provide Medical Equipment Maintenance and Repair

Provide medical equipment maintenance and repair of deployed medical equipment.

NO.	Scale	Measure
01	Yes/No	Nonavailability of medical equipment did not degrade, delay, or disrupt unit operations and medical treatment of wounded, injured, and sick Soldiers.
02	Time	To provide emergency repair of medical equipment in AO.
03	Time	To refine medical equipment maintenance and repair operations program for AO after receipt of warning order.
04	Number	Of medical equipment (each type) with remote prognosis/diagnostics capability in the AO.
05	Number	Difference between planned medical equipment maintenance and repair operations requirements and actual requirements in AO. Of repairs completed using remote prognostics/diagnostics in AO.
06	Number	Of medical equipment (each type) in AO with original equipment manufacturer (OEM) training required. planned medical equipment repair support performed in AO.
07	Number	Of medical equipment that requires repair per month. technicians OEM trained in AO.
08	Number	Of instances when medical capability was unavailable due to inability to repair medical equipment in AO.
09	Number	Of incidents involving medical equipment suspected of malfunctioning and causing further injury or death of a Soldier in the AO.
10	Number	Of incidents involving medical equipment that actually caused further injury or death of a Soldier in the AO, per The Safe Medical Device Act of 1990 and the Medical Device Amendments of 1992 (21 USC Section 360i (b) and 21 CFR part 803).
11	Number	Of repairs with remote prognostic/diagnostic capability which could not be remedied in the AO.

Supporting Collective Tasks:

Task No.	Title	Proponent	Echelon
-----------------	--------------	------------------	----------------